



BHASKARACHARYA COLLEGE OF APPLIED SCIENCES

(University of Delhi)

Sector – 2, Phase – 1, Dwarka, New Delhi – 110075, Phone- 011-25087597

Website:<http://www.bcas.du.ac.in>, Email: bhaskaracharya.college@gmail.com

E-Procurement Tender Notice

The College invites ONLINE bids as per Two bids System (Technical and Financial) from eligible bidders through e-procurement <https://eprocure.gov.in/eprocure/app>

Our Enquiry Ref No : BCAS/Quot./2018/ FT-NR/A.2/FH01
Dated : __20__/_02__/_2018__
Tender fee (Rs.) : 250/-
Bid download start Date and Time : __20__/_02__/_2018__, 1400hrs
Bid submission start Date and Time : __21__/_02__/_2018__, 1000 hrs
Last date and time for Bid Submission : __14__/_03__/_2018__, 1430 hrs
Date and Time of Bid Opening : __15__/_03__/_2018__, 1430 hrs
Tender Value (Rs.) : 110000/-
EMD(Rs.) : 2500/-
Bid Validity : Up to 31.03.2018

**Subject: Invitation of ONLINE 2 fold bids for the procurement of __ Fumehood __
by the Department of ____ Food Technology _____.**

Dear Bidder,

Only the online bids are invited for the procurement of Laboratory Equipment/____ Fumehood____
____ as per the details attached as Enclosure-I of the item(s).

S.No.	Name of Item(s)	Quantity Required	Minimum Specifications
1.	Fume Hood	1	Enclosure I
2.			

IMPORTANT:

- All details regarding the subject tender are available on websites www.bcas.du.ac.in and <https://eprocure.gov.in/eprocure/app>. Any change/ modification in the Tender Enquiry/ Tender Document will be intimated through above websites only. Bidders are therefore, requested to visit the websites regularly to keep themselves updated.
- Bids shall be submitted online only at CPPP website: <http://eprocure.gov.in/eprocure/app>
- Manual bids shall not be accepted.
- For submission of E-Bids, bidders are required to get themselves registered with <http://eprocure.gov.in/eprocure/app>
- Bidders are advised to follow the instructions provided in the 'Instructions to the Contractors/Bidder' for the e-submission of the bids online through the Central Public Procurement Portal for e Procurement at <https://eprocure.gov.in/eprocure/app>
- Bid documents may be scanned with minimum 100dpi with black and white option in pdf format.

It is required that the following instructions should be carefully followed including detailed terms and conditions attached overleaf as Annexure 'A', while submitting your offer; otherwise your offer may not be considered.

- All the communication with the college should be addressed only to "Principal, Bhaskaracharya College of Applied Sciences, Sector 2, Phase I, Dwarka, New Delhi- 110 075." (hereinafter called the Principal)
- Online Quotations will be two fold (a) one technical bid consisting of all technical details and supporting documents (b) another financial bid containing items wise price for the items mentioned in the technical bid. Bidders will not be permitted to alter or modify their bids after expiry of the deadline for receipt of bids.
- Financial bids of only those bidders will be opened and considered who qualify in their technical bid.
- Corrigendum, if any, will be published only on the above websites only.

Yours Sincerely,

Principal

Enclosure-I

Minimum Technical Specification of Fumehood

Working area size of chamber	910 x 910 x 610 mm (LxDxH).
Overall dimension including the base storage cabinet and main chamber	1210 x 1110 x 2400 mm.
Material of Construction	Galvanized iron sheets with epoxy painted
Inner area	The inner working area should be coated with teflon cloth sheets
Lower Base	Should be provided with the base storage cabinet double door.
Working platform	Black granite working table
Exhaust Unit	<ul style="list-style-type: none">➤ The exhaust unit should be fitted with heavy duty 1hp motor blower assembly to suck the fumes.➤ It should be fitted with a P-filter to prevent the environment from hazardous acid fumes. The P- filter (pre filter) should be washable type.
Door	The unit should have a vertical weight balanced sliding door made up of S.S/ Aluminium painted frame fitted with 6 mm toughened glass.
Electrical Fittings	Fitted with 4 nos. of switches and sockets in the inner area
Heating mantle	Three (500 ml capacity) with temperature controller
Others	<ol style="list-style-type: none">1. The hood should be fitted with 2 nos. of panel mounting remote fume hood valves 1 for water and 1 for gas. With outlet nozels fitted inside the working area.2. Fitted with LED tubes/ bulb3. Caster wheels4. Airflow controller with timer5. ISO and CE certified Parts and Product
Warranty	2 years
Installation at our place (Including all accessories)	Included in the price
Demonstration	Free of cost